

CJ



UNIVERSITY OF SOUTH CAROLINA

ADDENDUM NO. 1

TO: ALL VENDORS

FROM: Reed Bickers *[Signature]*

SUBJECT: Sealed Bid # USC-IFB-2467-RB
OPENING DATE AND TIME: REMAIN UNCHANGED

DATE: July 17, 2013

This Addendum No. 1 modifies the Bid invitation only in the manner and to the extent as stated herein.

Question #1: The specifications for the instrument in this bid are clear, however, there no critical options mentioned such as computer, autosampler, cooling device, etc. Please specify those important items.

Answer: Critical options for Inductively Coupled Plasma optical emission spectrometer: Yes/no

- 1. Chiller with a turbine pump and water recirculator. _____
- 2. Internal standards kit _____
- 3. Autosampler with 4 racks, standard rack, pump wash facility, probe. _____
- 4. Computer workstation with Windows 7 Professional with ICP software _____
- 5. Flat panel computer monitor _____
- 6. Laser Printer that has capability of networking _____
- 7. Sample introduction kit with spare parts _____
- 8. Additional peristaltic tubing _____
- 9. 1 full day of training after installation _____
- 10. Minimum of 3 days additional training on site by ICP training specialist _____

BIDDER SHALL ACKNOWLEDGE RECEIPT OF ADDENDUM NO. 1 IN THE SPACE PROVIDED BELOW
 AND RETURN IT **WITH THEIR SEALED BID RESPONSE**. FAILURE TO DO SO MAY SUBJECT BID TO REJECTION.

Authorized Signature

Firm

Date