



## AMENDMENT TO SOLICITATION

### Amendment No. 1

**DATE:** November 24, 2014

**TO:** ALL VENDORS

**FROM:** Mac Stiles, Procurement Manager

**SOLICITATION NUMBER:** USC-RFQ-2717-MS

**DESCRIPTION:** Provide and Deliver Laboratory Glove Box

---

This **Amendment No. 1** modifies the **Request For Quotations** only in the manner and to the extent as stated herein.

### **Vendor Questions**

1. The spec states "Contains four (4) ports for electrical connections/wiring to the outside of the box." Could you please elaborate? What type of connection are they looking for?

**We are looking for ports in which we can get wiring through and/or connect banana clips to, in order to get an electrical feed into the box. These are sometimes referred to as BNC feedthroughs.**

2. The spec states "Contains two (2) large (12-16") butyl gloves. This size does not make sense for butyl gloves. Butyl gloves come in either 8.5, 9 or 10. I am not sure what they are asking for.

**The box should come with a set of butyl gloves. We would like gloves with a port size of 8" or larger, a length of 12" or larger, and a thickness of 15ml or larger.**

3. There is no mention of BNC feedthroughs. Is this something that is needed? If so, how many?

**See answer to #1.**

4. There is no mention of internal charcoal solvent removal traps. Is this something that is needed?

**An internal charcoal solvent removal trap will be acceptable.**

---

Bidder shall acknowledge receipt of Amendment No. 1 in the space provided below and return it with their bid response. Failure to do so may subject bid to rejection.

---

Authorized Signature

---

Name of Offeror

---

Date