



AMENDMENT NO. 1 TO SOLICITATION

TO: ALL VENDORS

FROM: Ashley Kennedy-Shell, Procurement Officer

SUBJECT: SOLICITATION NUMBER: USC-RFQ-3508-AS

DESCRIPTION: Furnish and Deliver Mechanical Testing Equipment

DATE: July 2, 2019

This Amendment No. 1 modifies the Request for Qualifications only in the manner and to the extent as stated herein.

Vendor Questions and Answers. See the following pages.

BIDDER SHALL ACKNOWLEDGE RECEIPT OF AMENDMENT NO. 1 IN THE SPACE PROVIDED BELOW AND RETURN IT WITH THEIR BID RESPONSE. FAILURE TO DO SO MAY SUBJECT BID TO REJECTION.

Authorized Signature

Name of Offeror

Date

1. Can you tell me why is 17-4 specified for material on most of the fixtures?
Is A2, A4 or 316 a suitable substitute?

We were referencing what was specified as material used for manufacturing of each test fixture from our research. But if your alternative material selections are compliant to the same testing requirements for repeated usage, we are fine with your recommendations of A2, A4, or 316 steel as suitable substitutes.